CERTIFICATE OF CONFORMITY

with

IEC61000-4-30 :2003 Class A Testing and Measurement Techniques -Power Quality Measurement Methods

Certificate No:	TI/132/8407
Manufacturer:	Dranetz-BMI 1000 New Durham Road Edison, NJ 08818-4019 USA

Product Type:	Power Quality Measurement Instruments		
Model No:	Model PX5 Power Xplorer and variant models: PX5-400 Power Xplorer, 4400 Power Guide, 440S Power Guia, 440J Power Visa & 440 Power Visa.		
Report Ref:	61000-4-30 Technical Construction File – electronic version Document Number DTI-PQ-TC1	Date Created	7 th December 2004
	Amendment 1	Date Created	5 th January 2005

In our capacity as an independent product safety consulting firm, and an EU Competent Body for the EMC Directive and Notified Body for the Low Voltage Directive, from the evidence of an examination of the above report, we hereby certify that the product described above complies with the requirements of IEC 61000-4-30:2003 Class A Testing and Measurement Techniques - Power Quality Measurement Methods.

Issued By:	Technology International (Europe) Limited 60 Shrivenham Hundred Business Park Shrivenham, Swindon, SN6 8TY Tel: (+44) 1793 783137 Fax: (+44) 1793 782310		
Signed:	A. L. Brown		
Name:	A.L. Brown BSc, MSc, CPhys, MInstP	Date: 13 th January 2005	